

DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

REPORT N. 8172-20 Rev.00

Customer: **IDEANTIS SRL**
VIALE MONTE NERO 80 - 20135 - MILANO

TIME SCHEDULE

Acceptance N.: 20-7995
 Samples receiving date: 26/11/2020
 Start test date: 01/12/2020
 End test date: 01/12/2020
 Coronati Consulting Operator: Dr. F. Bergonzini

TEST LABORATORY

Coronati Consulting srl Via L. Gavioli, 3 I-41037 Mirandola (MO) Certified ISO 9001/ ISO 13485



REFERENCE DOCUMENTS

UNI EN 14683:2019 "Medical face masks - Requirements and test methods" - Annex C

TEST SAMPLE IDENTIFICATION

Name (1÷5): Maskèdra
 Sample Typology: Mascherina chirurgica
 Composition: poliuretano+poliammide12+Spunbond+Melt blown+natural rubber+poliestere
 Quantity tested: 5
 Code (REF): N/A
 Batch: 001
 Manufacturing date: ottobre 2020
 Expiry date: Non applicabile
 Sterilization Method: Not sterile
 Sterilization Batch: N/A
 Sterilization Date: N/A
 Sterilization Unit: N/A

The information concerning the test sample were provided by the Customer. All data related to the test sample are under the responsibility of the Customer and have not been verified by the test laboratory.

Issue Date	Rev.	Change Description	Prepared by: Dr. F. Bergonzini (Laboratory Technician)	Verified and Approved by: Dr. Renzo Giovanni Coronati (Managing Director Laboratory)
01/12/2020	00	First Issue		

This test report is digitally signed by Dr. Renzo Giovanni Coronati.
 The digital signature has legal value according to Italian D. Lgs. 82/2005 and subsequent amendments.

*The sampling is performed by the Customer. The test results are related only to the test samples as received.
 This report shall not be reproduced except in full without the written approval of Coronati Consulting srl.*

PROCEDURES

All procedures used during this study are recorded in the Laboratory Coronati Consulting s.r.l.

PURPOSE

The purpose of this procedure is to evaluate the suitability of the sample in meeting the breathability requirement as specified in the reference standard document. Determining the differential pressure between the two sides of the sample crossed by a constant and controlled air flow, this procedure allows you to evaluate whether this device has sufficient air permeability.

OPERATING METHODS (according to UNI EN 14683:2019 Annex C, par. C.4)

Each sample is placed in the holder and subjected to an 8 l/min air flow generated by a vacuum pump. The differential pressure on the tested area is read directly using a differential manometer. Tested areas number and location on sample are reported in the following figure 1.




Figure 1

ACCEPTANCE CRITERIA (according to UNI EN 14683:2019 par. 5.2.7, table 1)

	Type I ^(a)	Type II	Type IIR
Differential pressure	< 40 Pa/cm ²	< 40 Pa/cm ²	< 60 Pa/cm ²
(a) Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by health care professionals in an operating room or in other medical settings with similar requirements.			

RESULTS

N° test area	Differential Pressure (ΔP)				
	Flow rate: 8 L/min				
	Sample 1 (Pa/cm ²)	Sample 2 (Pa/cm ²)	Sample 3 (Pa/cm ²)	Sample 4 (Pa/cm ²)	Sample 5 (Pa/cm ²)
1	34,47	34,22	35,47	37,65	36,49
2	34,35	42,02	33,98	34,27	29,12
3	32,29	34,06	37,29	34,47	40,47
4	40,69	36,37	37,37	35,14	35,37
5	35,49	38,53	36,45	37,88	35,71
ΔP Average	35,46	37,04	36,11	35,88	35,43
MASK TYPE	DIFFERENTIAL PRESSURE (ΔP) IDONEITY				
I e II	PASS	PASS	PASS	PASS	PASS
IIR	PASS	PASS	PASS	PASS	PASS

ANNEXES

Annex 01: Sample Composition